

Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DRIVE 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 322534
Manufacturer RaySafe

CALIBRATION INFORMATION

As found Not performed
As left 2024-02-28
Adjustment done Yes
Tested by Ujjaini Joseph

Approved by Rasmus Brännström
Rasmus Brännström
Finalization operator
Certificate date 2024-03-11

CALIBRATION RESULT

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %, Air kerma rate: 5 %
Time: 0.5 %, Voltage: 2 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DRIVE 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 322534
Manufacturer RaySafe

CALIBRATION INFORMATION

As found Not performed
As left 2024-02-28
Adjustment done Yes
Tested by Ujjaini Joseph

Approved by

Rasmus Brännström

Rasmus Brännström
Finalization operator

Certificate date

2024-03-11

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate

Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DRIVE 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Light
Serial number 312807
Manufacturer RaySafe

CALIBRATION INFORMATION

As found Not performed
As left 2024-02-20
Adjustment done Yes
Tested by Ida Wiholm

Approved by

Rasmus Brännström

Rasmus Brännström
Finalization operator

Certificate date

2024-03-11

CALIBRATION RESULT

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Illuminance: 3 %

Luminance: 3 %



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
BLOSSOM DRIVE 1290
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Survey
Serial number 319724
Manufacturer RaySafe

CALIBRATION INFORMATION

As found Not performed
As left 2024-02-26
Adjustment done Yes
Tested by Ida Wiholm

Approved by

Rasmus Brännström

Rasmus Brännström
Finalization operator

Certificate date

2024-03-11

CALIBRATION RESULT

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 % (RQA), 10 % (N-series)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.

