

Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com


CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 347434
Manufacturer RaySafe

CALIBRATION INFORMATION

As found Not performed
As left 2026-01-15
Adjustment done Yes
Tested by Philip Kinnander

Approved by



Marie Panula
Service Technician

Certificate date

2026-03-02

CALIBRATION RESULT

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %, Air kerma rate: 5 %
Time: 0.5 %, Voltage: 2 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 R/F
Serial number 347434
Manufacturer RaySafe

CALIBRATION INFORMATION

As found Not performed
As left 2026-01-15
Adjustment done Yes
Tested by Philip Kinnander

Approved by



Marie Panula
Service Technician

Certificate date

2026-03-02

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com


CUSTOMER INSTRUMENT

Product X2 MAM
Serial number 202282
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2026-01-22
As left 2026-02-28
Adjustment done Yes
Tested by Kristoffer Petric
Ujjaini Joseph
Vera Zhang

Approved by



Marie Panula
Service Technician

Certificate date

2026-03-02

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air Kerma: 5 %
Voltage: 2 % or 0.5 kV (without paddle), 2 % or 0.7 kV (with paddle)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com


CUSTOMER INSTRUMENT

Product X2 MAM
Serial number 202282
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2026-01-22
As left 2026-02-28
Adjustment done Yes
Tested by Kristoffer Petric
Ujjaini Joseph
Vera Zhang

Approved by



Marie Panula
Service Technician

Certificate date

2026-03-02

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 CT
Serial number 338960
Manufacturer RaySafe

CALIBRATION INFORMATION

As found Not performed
As left 2026-01-14
Adjustment done Yes
Tested by Erik Strömberg

Approved by



Marie Panula
Service Technician

Certificate date

2026-03-02

CALIBRATION RESULT

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Survey
Serial number 230653
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2026-01-08
As left 2026-01-12
Adjustment done Yes
Tested by Hussein Mahamod
Ujjaini Joseph

Approved by



Marie Panula
Service Technician

Certificate date 2026-03-02

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Air kerma: 5 % (RQA), 10 % (N-series)

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.



Calibration Certificate



Customer
UPSTATE MEDICAL PHYSICS
ANTHONY DR 6536
14564 VICTOR
US

Laboratory
Unfors RaySafe AB
Björklundabacken 10
SE-436 57 Hovås
Sweden
+46 31 719 97 10 (phone)
+46 31 910 950 (fax)
customerservice.se@raysafe.com

CUSTOMER INSTRUMENT

Product X2 Light
Serial number 259179
Manufacturer RaySafe

CALIBRATION INFORMATION

As found 2026-01-21
As left 2026-02-27
Adjustment done Yes
Tested by Emmy Pantzar

Approved by



Marie Panula
Service Technician

Certificate date 2026-03-02

CALIBRATION RESULT

AS FOUND: All measured and tested values of this certificate were found to be in conformance with the specification.

AS LEFT: All measured and tested values of this certificate were found to be in conformance with the specification.

Relevant instrument specifications

Illuminance: 5 %

Luminance: 5 %

This laboratory is accredited according to ISO/IEC 17025:2017 by the "Swedish Board for Accreditation and Conformity Assessment" (SWEDAC) and the results shown in this certificate have been determined within the scope of accreditation unless stated otherwise in this certificate.

